

SPICES BOARD
KOCHI

Procurement of HPLC Systems for Quality Evaluation Laboratories of Spices Board on buyback

CORRIGENDUM FOR TENDER
Hindi Version Will Follow

17.12.2020

References:

Tender No. SB/QEL/MUMBAI/HPLC/2020 Dated: 06.10.2020
MSTCL Tender No. SPICEB/20-21/ET/3

Change in the technical specification.

Venders are requested to consider the following changes while submitting their bids for this tender

Pump	
3. Suitable for nano, micro and semi-micro separations	This specification is deleted
5. Resolution- 3 nanoliter per step	This specification is deleted
10. Pulsation 0.1 MPa	Pulsation < 0.2 MPa
Autosampler	
10. Operating pH range 1 - 14	Operating pH range 1-12 or better (Specify)
Degasser	
2. degassed flow-line capacity 380ul	This specification is deleted
Column Oven	
1. Temperature control from 4-85 °C	Temperature control from 5-85 °C
4. Storage capacity - minimum 3 columns	Storage capacity - minimum 2 columns
Fluorescence Detector	
2. Wavelength range - 200nm to 650nm	2. Wavelength range - 200nm to 650nm or better


SRILATHA C. M

Scientist - C (QEL in- charge)

