

# **SPICES BOARD**

## **TENDER NOTICE**

SB/QEL/PUR/MUM/LC MSMS/13

On behalf of the Spices Board, Ministry of Commerce, Govt. of India, Tenders are invited for the supply of following equipment conforming to the specifications and other requirements given in the Annexure-I.

---

### **LC MS/MS SYSTEM**

for

**Spices Board Regional Quality Evaluation Laboratory, Mumbai**

---

#### **Instructions to the Bidder:**

1. Preparation of Bids: The tenders should be submitted under the two bid system, viz., Technical bid and Financial bid.
2. Delivery of Tenders: The tender shall be submitted to the designated address either by Speed Post or by Courier, so as to reach our office before the due date and time. The details are mentioned in the Terms and Conditions, attached.
3. Opening of Tenders: The offers/bids will be opened by a Committee duly constituted for this purpose at the specified place, date and time in the presence of such tenderers or their representatives as may choose to attend. The technical bids will be opened first and they will be examined by a Technical Committee, which will decide the suitability as per the specification and requirement. The financial offer bid will be opened only for those offers/bids, which technically meet all the requirements and specifications. [The bidders, if interested, may also participate on the tender opening date and time]. The participating bidders should produce an Authorization Letter from their companies to participate in the tender opening day. Only one representative per tenderer will be allowed to participate [during the opening of tender].
4. Rejection of Tenders: The Spices Board reserves the right to reject any or all of the tenders received without assigning any reason.
5. Terms of Delivery: The quoted instrument(s) and all accessories and supporting instruments should be supplied, and the installation/commissioning of the same should be completed as specified in the Terms and Conditions mentioned. Signed copy of the Terms and Conditions should be submitted along with the Tender.

Secretary

Spices Board, Kochi-25

## **TERMS AND CONDITIONS OF THE TENDER**

Tender should be submitted in the prescribed tender document downloaded from the website [www.indianspices.com](http://www.indianspices.com). The tenders submitted should strictly follow the instructions specified below, failing any of which the tenders would be deemed incomplete and liable for rejection.

1. The last date for the submission of the tender will be at 5:00 pm on 11.11.2013

at the following Address:

The Deputy Director (Marketing)

Spices Board, Regional Office

First Floor, Banking Complex II, Sector 19-A

Vashi, Navi Mumbai-400 703. Ph: 022 2784 3093

Email: [sbromumbai@gmail.com](mailto:sbromumbai@gmail.com) / [sbqelmumbai@gmail.com](mailto:sbqelmumbai@gmail.com)

2. Tenders received after the due date and time for whatever reasons will not be considered. The offer/bids should be submitted in a manner of two bid system, i.e. [1] Technical Bid and [2] Financial Bid. Tenders received with both the bids in the same cover shall be summarily rejected.
  - a. The Technical Bid should consist of all technical details along with commercial terms and conditions. The bid should be submitted in the format given in Annexure-I (Technical Specifications), and should provide specific yes / no answer to each point in the specification. Additional details, if any, should be given in the Remarks column of the format. All claims of instrument performance as required in the specifications should be substantiated by adequate instrument readouts and data. The Technical Bid should be **in a sealed cover** superscribed "Technical Bid – LC MS/MS System, Mumbai, August 2013" and addressed to the Deputy Director Marketing (See 1 above).
  - b. Financial bid should indicate item wise price for the items mentioned in Technical Bid. Detailed pricing including all charges for accessories, supporting instruments, items, quoted in Indian Rupees, extended warranty, shipping, insurance, taxes, installation charges etc. should be clearly indicated. Unit rate along with split up para of the item quoted shall be given. The rates should be indicated clearly both in figures and words. If there is any variation between the rates in figures and in words, the rates given in words will be considered for evaluation. The rates should be firm and valid for 90 days from the date of opening of financial bid. The tenderer should put full signature atleast any scoring off/corrections/or overwritings. The rates quoted should not be subject to any price variation clause. A signed covering sheet should be attached to the financial bid in the format given in Annexure-II (Summary of financial bid). The Financial Bid should also be **in a sealed cover** superscribed "Financial Bid – LC MS/MS System, Mumbai, October 2013" and addressed to the Deputy Director Marketing (See 1 above).
  - c. Both the sealed covers viz., Technical and Financial Bids should be put into a big cover, along with EMD (See 3 below) and sealed. This cover should be superscribed "**Technical and Financial Bids for LC MS/MS System, Mumbai, October 2013**" and addressed to the Deputy Director Marketing (See 1 above).

- d. The tenders will be opened at the Regional Office of Spices Board, Mumbai on 11.11.2013 at 10:00 AM. Only the Technical Bids will be opened on this date. The bidders who qualify in the Technical Bid will be informed of the date of opening of the Financial Bids.
3. **EMD:** (a) The tenderer should submit an EMD equal to 2% of the total quoted price along with the tender by way of Demand Draft favouring “**Secretary, Spices Board**” and “**payable at Kochi**” from any nationalized bank valid for Six months . (b) Bank Guarantee will not be accepted towards EMD]. The Technical Bid [without EMD would be considered as UNRESPONSIVE and] will be rejected unless any category of tenderer is specifically exempted by the Government from payment of EMD for which reasons and proof have to be enclosed. (c) The EMD will be returned without any interest to the unsuccessful bidders after the award of the contract.

**Agreement Purchase Order for forfeiture of EMD:** EMD will be forfeited if the tenderer

- (i) withdraws his tender or backs out after acceptance
  - (ii) fails to remit the performance security
  - (iii) violates any of the conditions prescribed in the tender notice
  - (iv) revises any of the terms quoted, during the validity period
4. **Performance Security:** The successful bidder has to submit a Performance Security for an amount of 20% of the order value in the form of a **Bank Guarantee** from any nationalized bank and should be valid for 60 days beyond the date of completion of all contractual obligations including warranty obligation, and it will be released after 60 days beyond warranty period.
  5. If an Indian Agent is involved, the following documents must be enclosed:
    - a. Foreign principal’s proforma invoice indicating the Commission payable to the Indian Agent and the nature of after sales service to be rendered by the Indian Agent.
    - b. Copy of the agency agreement with the foreign principal and the precise relationship between them and their mutual interest in the business.
    - c. The enlistment of the Indian agent with Director General of Supplies & Disposals under the Compulsory Registration Scheme of Ministry of Finance.
  6. The offer/bids should be sent only for a machine that is available in the market and supplied to a number of customers. A list of customers in India with details should be furnished with the tenders in Sl.No.IV. Tenders for a prototype machine will not be accepted.
  7. Original catalogue (not any photocopy) of the quoted model duly signed by the principals must accompany the tender.
  8. In case the Technical Specification for the Instrument calls for Performance Data for the quoted model, the same has to be submitted completely in all respects in a booklet form for ready reference. This data has to be generated using the quoted model, by the bidder’s own application laboratory in India, as a proof of capability to provide application support. In case more than one model is quoted, the performance data for each model should be submitted separately. Generalized application notes or research papers will not be accepted as proof of performance, as these may not comply with the specific requirements of the Instrument. Spices Board reserves the right to witness either the reproduction of the data generated by the bidder’s laboratory or to analyze the required samples / solutions at the bidders laboratory in case the need arises.

9. **Delivery Schedule:** The bidder should indicate clearly the time required for delivery of the instrument. In case if there is any deviation in the delivery schedule, liquidated damages clause will be enforced or penalty for the delayed supply period will be levied.
10. **Risk Purchase Clause:** In the event of failure of supply of the item/equipment within the stipulated delivery schedule, the Board reserves the right to purchase the item/equipment from other sources on total risk of the supplier under Risk Purchase Clause.
11. **Payment:** No Advance payment will be made for Indigenous purchase. Payment shall be made only after satisfactory installation is agreed to wherever the installation is involved.
12. **Warranty/Guarantee:** The offer should clearly specify five years warranty including all the spares/accessories and consumables required for the smooth and continuous functioning of the equipment in the format given in Section III.
13. **User's List:** The offer should include a list of detailed users in the format given in Section IV.
14. **Late offer:** The offers received after the due date and time will not be considered.
15. **Acceptance and Rejection:** Spices Board reserves the right to accept whole or any part of any tender or portion of the quantity offered or reject it in full without assigning any reason.
16. **Disputes and Jurisdiction:** Any legal disputes arising out of any breach of contract pertaining to this tender shall be settled in the court of competent jurisdiction located within the city of Ernakulam, Kochi.
17. The tenderer must sign and affix seal on every page of the tender document and furnish a certificate acknowledgement as shown below:

Acknowledgement by the Tenderer

It is hereby acknowledged that we have gone through all the conditions mentioned above and we agree to abide by them.

SIGNATURE OF TENDERER

SEAL OF THE COMPANY

DATE

Annexure I

## Technical Specifications

(To be submitted by the Bidder)

---

**Instrument:** LC MS/MS System

**Quotation Reference:** SB/QEL/PUR/MUM/LC MSMS/13, dated 8-August-2013

**Location of Supply and Installation:**

Quality Evaluation Laboratory,  
Spices Board, (Government of India)  
Third Floor, Banking Complex-II, Sector 19-A  
Vashi, Navi Mumbai-400703 (Ph: 022-2784 3093)

*The technical bid is to be submitted in this form, duly signed by the authorized signatory. Any incompleteness in filling up this form or failure to attach the required supporting data or documents, may result in rejection of the quotation.*

---

### Section I: Details of the Quotation

Quotation Submitted by	
Quotation Reference No.	
Date of Submission of Quotation	
Details of EMD submitted (if any)	
Full Address	
Name of Principal Supplier (If not self)	
Authorized by the Principal Supplier in India (attach details)	
Authorized Contact Person	
Designation	
Contact Details	
List of supporting documents submitted along with the quotation:	

## Section II: Compliance with Technical Specifications

### LC MS/MS System

(In case more than one model / combination is quoted, attach separate compliance sheets for each)

All items required for the full-fledged function of an advanced, high sensitive LC MS/MS system, including all software, accessories and supporting instruments as outlined below are to be quoted.

**Against each point in the technical specification, compliance must be marked as YES / NO**

**A. HPLC SYSTEM.** All components of the HPLC system should be from the same supplier, and should amenable to control / data acquisition from the MS Software.

Quoted Model:

	Required Technical Specifications	Complying? (Yes/No)	Remarks	Documentary Evidence enclosed?
1	Quaternary Gradient Pumping System capable of providing precise and controlled flow suitable for an MS/MS system.			-
2	<i>Purging of pumps, manually as well as through software</i>			-
3	<i>Suitable for operation as front-end for MS/MS System, compatible with the software controlling the MS/MS System</i>			-
4	<i>Flow rate setting: 0.01ml/min to 10.0 ml/min or less</i>			-
5	<i>Resolution : 3 nanoliter per step</i>			-
6	<i>Flow rate accuracy: 1% minimum</i>			-
7	<i>Flow rate precision Maximum 0.3% RSD</i>			-

8	<i>Gradient curve selections</i>			-
9	<i>Maximum discharge pressure at least 40 M.Pa</i>			-
10	<i>Pulsation 0.1 M.Pa</i>			-
11	<i>Safety functions like leak sensor, high pressure and low pressure limits</i>			-
<b>Autosampler</b>				
1	<i>Automated operation controllable through MS/MS Software</i>			-
2	<i>Automated autosampler purging through software</i>			-
3	<i>Total volume sample injection, variable injection volume</i>			-
4	<i>Injection volume setting range 0.1ul to 100ul</i>			-
5	<i>Vials capacity at least 1.5ml</i>			-
6	<i>Number of vials - at least 100</i>			-

7	<i>Injection volume accuracy - 0.1% error maximum</i>			-
8	<i>Injection volume precision - 0.3% RSD maximum, for 10ul injection</i>			-
9	<i>cross-contamination limit - 0.005% maximum</i>			-
10	<i>Number of repeat injections - 30 / sample</i>			-
11	<i>Operating pH range 1 - 12</i>			-
12	<i>Rinsing needle - settable through software, before and after injection</i>			-
13	<i>Operating temperature range 4 - 35°C</i>			-
<b>Degasser</b>				
1	<i>Online membrane type degasser</i>			-
2	<i>Minimum 5 lines - 4 mobile phases, 1 autosampler rinsing</i>			-
<b>Column Oven</b>				



1	<i>Temperature control from ambient to 85°C</i>			-
2	<i>Temperature control precision 0.1°C max</i>			-
3	<i>Controllable through MS/MS software</i>			-
4	<i>Storage capacity - minimum 3 columns</i>			-
5	<i>Safety functions like leak sensor, high temperature cut-off</i>			-
<b>Columns</b>				
1	<i>The following high resolution, end capped columns suitable for analysis of azo dyes and natural products. The columns should be quoted from any of the following two suppliers : Phenomenex (Luna series), Waters (Symmetry series)</i>			-
	<i>C-18 5cm x 4.6mm x 5um - 10 Nos.</i>			-
	<i>C-18 10cm x 4.6mm x 5um - 5 Nos.</i>			-
	<i>C-18 25cm x 4.6mm x 5um - 2 No.</i>			-
	<i>C-8 5cm x 4.6mm x 5um - 5 Nos.</i>			-

	<i>C-8 10cm x 4.6mm x 5um - 5 Nos.</i>			-
	<i>C-8 25cm x 4.6mm x 5um - 1No.</i>			-
	<i>Amino 25cm x 4.6mm x 5um - 2No.</i>			
2	<i>The columns should be supplied with certificates citing batch no, particle size distribution, inertness, metal sensitivity, hydrophobic index and column performance checks</i>			-
3	<i>20 Nos. each of C-8 and C-18 guard columns, from the same supplier as the columns, should be provided. Replaceable cartridge-type guard columns are desirable. In this case, two separate cartridge holders are to be quoted.</i>			-

**B. MS/MS SYSTEM. A state-of-the-art, high sensitivity triple quadrupole mass spectrometer.**

Quoted Model:

	<b>Required Technical Specifications</b>	<b>Complying? (Yes/No)</b>	<b>Remarks</b>	<b>Documentary Evidence enclosed?</b>
<b>Ion Sources</b>				
1	<i>Dual / Orthogonal or off-axis spray or any other equally efficient technology capable of avoiding interference from solvents and other extraneous matter</i>			-
2	<i>ESI and APCI probes to be provided, with facility of interchanging easily by the user, and auto-detection of installed probe by the instrument and software.</i>			-
3	<i>The source should have self-contained heating system for achieving desired operational temperatures.</i>			-
4	<i>Sprayer position should be adjustable in vertical and horizontal directions with adjustment scales for easy optimization.</i>			-
5	<i>Facility for visual examination of spray should be available.</i>			-
6	<i>The source should be easily removable from the system to facilitate user cleaning <u>without venting the vaccum</u>, with automatic shutdown of system while the source / probe</i>			-

	<i>is being removed.</i>			
7	<i>The source should be able to handle 0.5ul/min to 2ml / min flow, without flow splitting.</i>			-
8	<i>The source housing should be fully vented to avoid contaminating laboratory air.</i>			-
9	<i>All source parameters to be adjustable through software.</i>			-
<b>Vacuum System</b>				
1	<i>A differentially pumped vacuum system, including air-cooled turbomolecular pump, capable of supplying the required vacuum to the instrument. The system should have vacuum safety features to prevent damage to the instrument in case of failure.</i>			-
2	<i>All accessories required for the proper functioning of the vacuum system should be supplied.</i>			-
<b>Mass Analyzer</b>				
1	<i>Triple quadrupole MS/MS System</i>			-
2	<i>Mass Range at least 50 - 1500 AMU</i>			-
3	<i>Mass Accuracy 0.01% over the entire mass range</i>			-
4	<i>Mass resolution - unit resolution over the entire mass range. Higher resolution capability desirable.</i>			-
5	<i>Mass Stability: 0.1 dalton over 24 hours (please provide graphical data)</i>			<input type="checkbox"/>
6	<i>Collision cell - high pressure collision cell for high sensitivity MRM studies without loss in sensitivity</i>			-
7	<i>Scan Types:</i>			
	<i>Full Scan MS</i>			
	<i>Selected Ion Monitoring in both Q1 and Q3</i>			
	<i>Precursor Ion Scan</i>			
	<i>Product Ion Scan</i>			

	<i>Neutral Loss Scan</i>			
	<i>MRM</i>			
8	<i>Scan Speed of at least 10,000 daltons / second or above, with user selectable scan speed.</i>			-
9	<i>Minimum MRM dwell time 1 millisecond</i>			-
10	<i>Minimum Number of MRMs: 200 MRMs in a single time period, with no loss in sensitivity for co-eluting components at any one point of time.</i>			-
11	<i>No significant crosstalk should be detectable for reserpine at the following conditions : Infusion of 0.2 pg/ul, for MRMs 609/195 and 100/195 at 2 milliseconds dwelltime (please provide graphical data)</i>			<input type="checkbox"/>
12	<i>+ve / -ve polarity switching time between alternate MRM scans: 50ms or less</i>			-
13	<i>A high sensitivity, high throughput detector with zero dead time, low noise and high accuracy at low level detections.</i>			-

### C. SOFTWARE

Name and Version:

	<i>Required Technical Specifications</i>	<i>Complying? (Yes/No)</i>	<i>Remarks</i>	<i>Documentary Evidence enclosed?</i>
1	<i>Latest model computer with DVD writer and laser printer, suitable for operation with the software, should be quoted.</i>			-
2	<i>The software should be user friendly, capable of controlling the entire functions of HPLC and MS/MS, with required licenses for system suitability calculations; all accessories required for its operation should be quoted.</i>			-
3	<i>The software should have full-fledged capability of calculating system suitability parameters, both for MS and HPLC.</i>			-
4	<i>Full-fledged reporting facility with capability of customized user-designed report formats should be available.</i>			-
5	<i>The software should necessarily have the following features:</i>			-
a	<i>Retention-time based programming / monitoring of MRM scans, so that only desired MRMs are monitored in desired retention time ranges</i>			-
b	<i>Fragment interpretation tools</i>			-
c	<i>Automated quantitative data processing features</i>			-
d	<i>Option for naming MRM transitions</i>			-

--	--	--	--	--

**D. SYSTEM PERFORMANCE REQUIREMENTS:**

The requirements in this section pertain to the working requirement of the system. Any claims made in the compliance statement should be substantiated by giving suitable detailed outputs from the quoted model of the instrument generated at the applications lab of the supplier, in the form of reports for easy reference. The reports should necessarily include the instrument output data, graphs and chromatograms. The claims in the compliance statement should be demonstrable at the time of installation. **The submitted data will need to be reproduced in the presence of Spices Board Officials, if deemed necessary.**

	<i>Required Technical Specifications</i>	<i>Complying? (Yes/No)</i>	<i>Remarks</i>	<i>Documentary Evidence enclosed?</i>
1	<i>Sensitivity of the instrument: In ESI +ve, MRM transition 609<sup>+</sup>195 at unit resolution (0.7± 0.1 amu), 1pg on-column HPLC injection of Reserpine should give a S/N ratio of at least 5000:1. For ten injections, %RSD should be &lt;5%. Chromatograms to be provided, with details of mobile phase, column and injection volume used.</i>			<input type="checkbox"/>
2	<i>Analysis of Mycotoxins: A limit of quantitation of Aflatoxin B, B2, G1 and G2 in chilli matrix of at least 1ppb <u>without using immunoaffinity column cleanup</u> in MRM mode (please provide full method details and chromatograms. Also mention the lowest LOQ achievable.)</i>			<input type="checkbox"/>
3	<i>Analysis of Illegal Dyes: A limit of quantitation of Para Red in chilli matrix of at least 0.5 ppb in MRM mode (please provide full method details and chromatograms. Also mention the lowest LOQ achievable.)</i>			<input type="checkbox"/>
4	<i>Analysis of Pesticide Residues: A limit of quantitation of Methyl Parathion in chilli matrix of at least 0.5ppb in MRM mode (please provide full method details and chromatograms. Also mention the lowest LOQ achievable.)</i>			<input type="checkbox"/>
5	<i>Pre-configured and optimized methods, with reference standards, for analysis of Pesticide Residues and Mycotoxins (provide details)</i>			-
6	<i>MRM List and product ion spectra for a wide range of compounds including pesticide residues, mycotoxins and synthetic dyes for screening and quantitation (provide details)</i>			-

**E. ACCESSORIES:** *All accessories required for the trouble free installation, demonstration and operation of the system should be quoted.*

	<i>Required Technical Specifications</i>	<i>Complying? (Yes/No)</i>	<i>Remarks</i>	<i>Documentary Evidence enclosed?</i>
<b>Gas Generators</b>				
1	<i>Gas Generators capable of supplying <u>all necessary gases</u> with the required purity, pressure and flow rate, as required for the LC MS/MS instruments should be provided. It should be complete with all necessary accessories.</i>			-
2	<i>If Gas Generators for a required gas is not available, two cylinders each of all required gases for operation of the system should be provided, to be quoted from local sources for supply at the installation site.</i>			-
3	<i>In case external compressors are required for the gas generators, the same should be portable, noiseless and highly durable.</i>			-
<b>Spares and Accessories</b>				

1	<i>All required standards of mass calibration and tuning, and any other material required to make the instrument functional should be provided</i>			-
2	<i>Pack of 200 vials, complete with septa and cap, should be included</i>			-
3	<i>An ultrasonic bath suitable for cleaning parts of HPLC and LC MS/MS should be included.</i>			-
<b>UPS</b>				
1	<i>A UPS System of <b>Numeric / Aplab Make</b>, with <b>Exide 6EL Tubular Batteries</b>, of appropriate capacity (minimum 10 KVA) and input/output phases as per requirement of the LC MS/MS System, capable of providing a <u>backup time of minimum 4 hours</u>, with comprehensive warranty of minimum 5 years inclusive of battery, should be quoted.</i>			-

**F. TERMS AND CONDITIONS:**

	<i>Required Technical Specifications</i>	<i>Complying? (Yes/No)</i>	<i>Remarks</i>	<i>Documentary Evidence enclosed?</i>
1	<i>All items should be quoted CIF Mumbai. All items, including accessories for the instruments should be delivered and installed at the following site of installation: <b>Spices Board Quality Evaluation Laboratory, Third Floor, Banking Complex II, Sector 19A, Vashi, Navi Mumbai, Maharashtra-400 703</b></i>			-
3	<i>A comprehensive warranty for 5 years for the <u>complete equipment</u> including independent accessories like UPS and gas generators</i>			-

	<i>should be provided. The warranty should include mass calibration performed yearly with valid PPG solution provided by the vendor. In case of breakdown, service should be available onsite within 24 hours (See Section II below).</i>			
4	<i>The quoted model should have a minimum of <u>5 successful installations</u> in India. The Complete users list for the <u>quoted model</u> in India, with contact addresses, emails and phone numbers should be provided (see Section IIIb below).</i>			<input type="checkbox"/>

Declaration by the Bidder

It is hereby declared that, the details given above are correct as pertaining to the quoted model, and all data and reports required for substantiation technical claims have been submitted.

SIGNATURE OF TENDERER

SEAL OF THE COMPANY

DATE



**Section III: Comprehensive Warranty**

**LC MS/MS System**

In case more than one model / combination is quoted, attach separate sheets for each.

Comprehensive Warranty for Instrument Model : \_\_\_\_\_

	<b>Required Technical Specifications</b>	<b>Complying? (Yes/No)</b>	<b>Remarks</b>
	<i>The warranty period should be 5 Years All breakdown of the instrument should be addressed within 24 hours by an onsite visit.</i>		
	<i>All parts of the instrument, accessories and supporting instruments should be covered under the warranty. If any accessory / spare part / consumable is <u>not covered under the comprehensive warranty</u>, then the itemised list of such items, including their prices during the period 2013-14 and average lifetime of the items should be provided.</i>		
	<i>The equipment should be installed at the proposed site, demonstrate the required objective of the test with the claimed limit of detection/quantification with the sample provided.</i>		
	<i>Training should be given to the staff at the site, for the operation of the instrument on demand</i>		
	<i>Any other Accessories /requirements like UPS, Gas cylinders etc.should be quoted and supplied with the instrument</i>		
	<i>Agency commission: Agency commission if any will be paid to the Indian agent in rupees on receipt of the equipment and after satisfactory installation.</i>		
	<i>Agency commission will not be paid in foreign currency under any circumstances. The details should be explicitly shown in tender even in case of Nil commission. The tender should indicate the percentage of Agency commission to be paid to the Indian agent.</i>		

## Section IV: Users List of the Quoted Model in India

### LC MS/MS System

In case more than one model / combination is quoted, attach separate sheets for each. Additional copies of this page may be used in case the number of users is large.

Users List for MS/MS Model : \_\_\_\_\_

\*Please mention the Operational Status as one of the following: Installed Only / In Operation. Do not include instruments that are not installed or not operational as of the date of opening of the tender.

User Name & Address	Date of Installation	Current Operational Status of the Instrument *	Area of Analysis	Name of the Contact Person	Designation	Contact Number	Email id

## Annexure II: Summary of Financial Bid

### LC MS/MS System

This sheet is to be attached as a covering sheet along with the formal technical bid of the quoted instrument. In case more than one instrument / model is quoted, separate sheets have to be attached with each quotation.

Financial Summary for the Instrument model : \_\_\_\_\_

Price Description		Price	Remarks (if any)
Quoted Currency			
Total Instrument Cost	[a]		
Total cost of Supporting Instruments	[b]		
Comprehensive Warranty Charges	[c]		
Any other charges	[d]		
Mumbai [a] + [b] + [c] + [d] =	[e]		
Conversion Factor to INR*	[f]		
Total Price in INR [e] x [f] =	[g]		
**Toal cost of Items quoted in INR =	[h]		
<b>Total cost of the Instrument [g]+[h]</b>	<b>[i]</b>		

\* Use the conversion factor as on the date of submission of the quotation.

\*\* Give full list of items quoted in Indian Rupees separately.

#### Declaration by the Tenderer

It is hereby declared that the details given above are correct as pertaining to the quoted model. The formal financial bid is enclosed including all relevant details.

SIGNATURE OF TENDERER

SEAL OF THE COMPANY

