



स्पाइसेस बोर्ड
भारत

**Indian Cardamom Research Institute
SPICES BOARD**

(Government of India)
(Ministry of Commerce and Industry)

Myladumpara
Kailasanadu P.O.,
Idukki District - 685 553
Kerala

भारतीय इलायची अनुसंधान संस्थान
स्पाइसेस बोर्ड
(भारत सरकार)
(वाणिज्य एवं उद्योग मंत्रालय)
मैलाडुंपारा,
कैलासनाड पी.ओ.
इडुक्की जिला - 685 553
केरल

No.Res/Path/DF/1/2011/
5.09.2011

QUOTATION NOTICE

Sub : Purchase of Spore Traps – Inviting Quotations – Reg.

Sealed competitive quotations are invited from suppliers of reputed brands for the supply of Spore Traps as per the specifications mentioned in the Annexure I.

The proposal for supply of equipment with all necessary accessories and spares, as detailed in the specification, should give break up cost, excise duty/sales tax/VAT, packing and forwarding charges, freight charges, insurances etc (as applicable) along with brochure and terms and conditions. The quotation should be superscribed as "Quotation for Lab equipments, Due by 26.09.2011" and sent to the Director (Research), Indian Cardamom Research Institute, Spices Board, Myladumpara, Kailasanadu P.O., Idukki Dist, Kerala. PIN 685553.

Late quotations will not be accepted.

Last date and time for the receipt of quotations (26.09.2011)

Date and time of opening of the quotations (27.09.2011)

Quotation should mention guarantee/warranty periods, after sales service details, delivery period, the customer list including name of contact persons with telephone no/email id etc. Onsite training personnel, if required, is to be conducted immediately after installation of the equipments. The quotations should be submitted at its lowest rate along with statement of compliance to the annexure I. Clarification if any can be had from Scientist, Plant Pathology Division, ICRI, Spices Board, Myladumpara. Rate for 5 years comprehensive warranty may also be quoted separately.

As Indian Cardamom Research Institute is registered with the Department of Scientific and Industrial Research (DSIR) for the purposes of availing customs duty exemption in terms of Government Notification No. 51/96- Customs dated 23 July, 1996 and Excise duty exemption in terms of Government Notification No. 10/97-Central Excise dated 1 March, 1997 as amended from time to time. So the price may be quoted in foreign currencies also. (Ref: No. TU/V/RG-DCE (520)/2010 dated 11.10.2010)

The Director (Research), reserves the right to reject any or all quotation without assigning any reasons thereof.

M.R. R...
Director (Research)

ANNEXURE I



Specification for 7 day recording volumetric spore trap

A compact unit with built-in pump, designed to sample airborne particles, such as fungus spores and pollens, continuously for periods of up to seven days without attention. Standard orifice size 2 x 14 mm. Alternative orifice size ½ x 14 mm Voltages: 230-250 volts 50 Hz; 110-115 volts 60 Hz; 12 volts d.c.

Overall height: 94 cm (37 inch)

Working area of trap with standard vane: 0.89 m² (9.61 feet²) through 53 cm (21 inch) radius.

Throughput: 10 l/min (0.35 cfm)

Drum speed: 1 revolution in 7 days at 2 mm/hour.

Standard orifice: 2 mm x 14 mm

Tape: Melinex clear tape 200 gauge.

Optional extras:

42/20 'Gelvatol' available in ½ kg packs; Alternative orifice, size 2 x 14 mm reducing to 5 mm; Additional drums; Alternative lid assembly for 24 hour sampling directly onto glass slide; Flowmeter for measuring throughput at the sampling orifice; Anodised aluminium carrying case with lid.

M.R. de... 1/9/11