

No.SB/BT/PUR/2007

Dated 25.02.2008

Sir,

Sub:- Inviting quotations for instrument to Spices Board-reg

Spices Board desires to purchase the following equipment. You are requested to send your lowest quotation for the instrument to us on or before 10th March 2008.

Instrument

Specification

Elisa plate reader

Provides quick, accurate & reproducible results. User Friendly menu system for convenient operation Dynamic system with built in stabilizer calculates results automatically, large nonvolatile memory stores more than 30 user entered tests, Supports multiple tests on n single plate. Automatic lamp saving mode capable of sorting, deleting and recalling tests extensive software for cut off mode view & prints reports simultaneously
Linear Measurement Range Photometric Accuracy:
0.0 to 3,000 absorbance units (A) +/-2% or +/-0.007 whichever is higher, from 0 to 1.5A, +/-3% from 1.5A to 3.0 A Optical measurement: Photodiode

You may also quote the price preferably in Indian rupees, guarantee period, terms and conditions and service contract etc. You are requested to send your lowest quotation in a **sealed cover superscribed QUOTATION FOR ELISA PLATE READER TO BIOTECHNOLOGY LAB, SPICES BOARD with in due date 10th March 2008.** addressed to The Director Research, Biotechnology Division, Spices Board, Sugandha Bhavan, N.H.Bye pass, palarivattom, Kochi-682 025. The Spices Board has the right to accept or reject the quotations and purchases will be as per the purchase formalities of the Spices Board.

Yours faithfully,

Director (Research)